

#### MODEL NUMBER DOC NO. PERFORMANCE SPECIFICATION 3224A1 PS3224A1 IEPE ACCELEROMETER REV L, ECN 15258, 08/02/19

mV/ m/s<sup>2</sup>



- MINIATURE SIZE
- HIGH SENSITIVITY
- EPOXY SEALED CABLE CONNECTION

PHYSICAL Weight, Less Cable
Mounting Provision
Connector [2]
Case Material
Sensing Element

### PERFORMANCE

Sensitivity [3] [4] Range F.S. For ± 5 Volts Output Frequency Response, ±10% Resonant Frequency, NOM Equivalent Electrical Noise Floor Amplitude Non-Linearity [5] Transverse Sensitivity, Max. Strain Sensitivity @250με

#### ENVIRONMENTAL

Maximum Vibration Maximum Shock Temperature Range Magnetic Sensitivity at 100 Gauss

## ELECTRICAL

Supply Current [6]

Compliance Voltage Range Output Impedance TYP Bias Voltage Discharge Time Constant Output Signal Polarity for Acceleration Toward Top Case Grounding

ENGLISH
0.007
[1]
Integral cable
Titanium
Planar Shear Piezoceramic

9 to 14	mV/g	
± 500	g	
0.3 to 20000	Hz	
>95	kHz	
0.023	g rms	
± 1%	%F.S	
5%		
0.14	g/με	

	_
900	grms
3000	±gpk
-60 to +300	°F
Laser welded, Epoxy Sealed Cable	
0.000009	g/Gauss
•	_

2 to 20	mA
18 to +30	V
100	Ω
+8 to +10	VDC
0.9 to 2.5	sec
Positive	
[7]	

SI	1
	l
0.2	grams
[1]	
Integral cable	
Titanium	
Planar Shear Piezoceramic	

± 4905	m/s <sup>2</sup>
0.3 to 20000	Hz
>95	kHz
0.225	m/s <sup>2</sup> rms
± 1%	%F.S
5%	
1.37	m/s²/με

9 to 1.4

900	grms	8820	m/s <sup>2</sup> rms
3000	±gpk	29400	± m/s² pk
-60 to +300	°F	-51 to 149	°C
elded, Epoxy Sealed Cable		Laser welded, Epoxy Sealed Cable	
0.000009	g/Gauss	0.8829	m/s <sup>2</sup> /T
	-	•	
2 to 20	mA	2 to 20	mA
10 to 120	\/	10 to ±20	1/

2 to 20	mA
18 to +30	V
100	Ω
+8 to +10	VDC
0.9 to 2.5	sec
Positive	
[7]	

# This family also includes:

Timo ranning ar	oo molaaco.				
Model	Sensitivity (mV/g)	Range (Gpeak)	Resolution (Grms)	Oper. Temp(°F)	TC (sec)
3224A2	2 to 5	2500	0.03	-60 to 300	0.3 to 1.0
3224A3	5 to 10	1000	0.037	-60 to 300	1.1 to 2.5
3224A4	1 to 2.5	5000	0.08	-60 to 300	0.3 to 1.0


Please, refer to the performance specifications of the products in this family for detailed description.

### Supplied Accessories:

- 1) Model 6725 installation removal wrench.
- 2) Sample of " Small Petro Mounting Wax".
- 3) Accredited calibration certificate (ISO 17025)

#### Notes:

- [1] Flat mounting surface for adhesive mount.
- [2] 10-32 Jack at end of integral 3FT.cable
- [3] Measured at 100 Hz, 1 g rms per ISA RP 37.2.
- [4] Actual sensitivity is given on a calibration certificate traceable to ISO 17025, supplied with each instrument
- [5] Measured using zero-based best straight-line method.
- [6] Do not apply power to this device without current limiting, 20 mA MAX.
- To do so will destroy the integral IC amplifier.
- [7] Case is grounded to electrical signal/power ground.
- [8] In the interest of constant product improvement, we reserve the rights to change the specifications without notice. It is the customer's responsibility to validate that a particular product with the properties described in the product specification is suitable for use in a particular application. Parameters provided in datasheets and / or specifications may vary in different applications and performance may vary overtime. All operating parameters, including typical parameters, must be validated for each customer application by the customer's technical experts.





21592 Marilla Street, Chatsworth, California 91311 Phone: 818.700.7818 Fax:818.700.7880 www.dytran.com For permission to reprint this content, please contact info@dytran.com